



Appendix I

Landfill Gas Monitoring Equipment Calibration Certificate



香港新界葵涌葵昌路58-70號永祥工業大廈10樓B室

Unit B, 10/F., Wing Cheung Industrial Building, 58-70 Kwai Cheong Road, Kwai Chung, New Territories, HK Tel: (852) 2751 7770 Fax: (852) 2756 2051 E-mail: rotter@rotter.com.hk

Calibration Report - Gas Detector

Calibration Report - Cas Detector						
PGM-2500 (QRAE III) LEL/O2/CO/H2S						
UNIT INFORMATI Customer: Penta Ocea		Serial # : M02A00 Firmware : V2. Cal date : 28-Jul-	14 Sensor:	QRAE III LEL/O2/CO/H2S Teddy		
SENSOR DATA:						
	LEL sensor (ME)	O2 sensor	CO sensor (Tox1)	H2S sensor (Tox2)		
Calibration dates:	28-Jul-2023	28-Jul-2023	28-Jul-2023	28-Jul-2023		
After Calibration levels	51%	18.10%	50 ppm	10.0 ppm		
Alarm levels (Low):	10.00%	19.50%	35 ppm	10 ppm		
Alarm levels (High):	20.00%	23.50%	200 ppm	20 ppm		
TWA Level:	(##)		35 ppm	10 ppm		
STEL Level:			100 ppm	15 ppm		
Status: Pump Speed Clock LEL Gas Selection LEL Calibration Gas LEL Custom Gas	Low Yes Methane LEL_custom_gas	Back Light Measure LEL measurement Gas LEL Custom Factor	Manual Average Methane 1.0			
	as Mix: (18% O2, 50ppm (Gas lot :302-402538759-74		
*** Fresh Air Calibrat	ion is highly recommended	d to proceed prior for mea	asurement each time.			
Replaced Parts:						
Notes:						
	l and checked under good	working condition				
The unit was campiated	and checked under good	Working Condition				
**Next calibration due	on or before 27 July 2024					

Serviced by Teddy Warg Rotter International Ltd

PROMAT (HK) LTD

寶時(香港)有限公司

901 New Trend Centre, 704 Prince Edward Road East, San Po Kong, Kowloon, HK Tel.: 2661 2392 Fax.: 2661 2086 email : sales@promat.hk http://www.promat.hk/



VERIFICATION CERTIFICATE OF CO2 Analyzer

Report No.

: 22040

Date

: 17/11/2022

Client

: Penta Ocean Concentric JV

EQUIPMENT TO BE VERIFIED

Equipment Name

: CO2 Analyzer

Supplier

: TES

Model No.

: 1307H

Serial No.

Date of Verification

: 200901259 : 17/11/2022

Due Verification

: 16/11/2023

VERIFICATION DEVICES USED

Reference Equipment

: CO2 in N2

CO2 in N2

Supplier

: NorLab

NorLab

Model No.

: H1013500PN

H1013.3VN

Lot#

: 0-353-790

1-006-27

Expiry date

: 12/2/2024

10/3/2025

Accuracy

: Within +/-2%

Within +/-2%

ENVIRONMENTAL CONDITION

Ambient Temp

: 25°C

Relative Humidity

: 57%

Verification Result

Test Number	Concentration (Mole%)	Results	
Test 1	500ppm	505ppm	
Test 2	0.50%	0.52%	

Remarks

- 1 The Gas reference used in this verification has traceable accuracy to Manufacturer Standard
- 2 The above equipment was operated by the competent person
- 3 Promat is Registered ISO9001:2015 Quality Management System in Sales, Repair and Calibration Services

Certification

Verification by

Checked by

Mr. Hei Kong / Technical Engineer

Ms. Ning Lee / Service Coordinator

~ I.	1	1.60	
(211	hration	Certificate	I
Vall	viauvii	Cu uncau	,

- de-manufacture excess				anticioner consended Communication and American State Communication Comm		
Cert	. Ref. No.: BW/XT/3	3RD/17974 Da	ate: 2022 09 0	2		
Customer: Renopipe Construction Company Limited			Purchase Order No.: SME-C-20-21-6/2020-76554			
九龍觀塘海濱道133	號萬兆豐中心6樓K2室		ete Cooo loz la	TANKOICE NO	. AB	
				7 INVOICE NO	: AP	
			mail: damonhuang@re	enopipe.com.hk		
Attn: Damon Huang	Attn: Damon Huang Tel: 3998 3193			Fax: 3998 3225 Mobile Phone		
User Details:						
Gas Detector Model: XT-XWH Calibration Record:	-IM-Y-OR Seri	ial No.: MA220-0127	09 Pump	S/N : 420373		
Inpectio	n before calibration	Ì	Visual inspe	ction	Functional Test	
Basic Unit - Case, Clip & Dis	play etc.		OK		OK	
Battery and charge etc.			OK		OK	
Motorized Pump			OK		OK	
Other items						
Type of Sensor		· · · · · · · · · · · · · · · · · · ·	Expiry Da	ite		
Oxygen Sensor						
CO & H2S Sensor						
Combustible(LEL) Sensor						
- C III (I			00/	00 (%)	LEL (%)	
Type of calibration	Date of calibration	H2S (ppm)	CO (ppm)	O2 (%)	LEL (70)	
3rd Calibration	2022 09 02	25	100	18	50	
Result of Calibration		ок	ОК	ок	ок	
Calibration Cost: (As per a	ttached invoice)	F.O.C			•	
Calibration remarks: Oxygen s	ensor replaced by new on-	e				
Warranty	: Oxygen Sensor 1 years v	warranty				
Next calibration date of	of this instrumer	nt will be: 20	09	02		
	IMPORTANT N	OTES TO RW C	AS DETECTOR	HSEDS		
USERS MUST READ THE					S EQUIPMENT	
AND FOLLOW THEIR OV						
All and detection instrument	4-4: 4114		libustica to commet	-1	Calibration is only	
All gas detection instrumentation on the market requires periodic calibration to accurately measure gas. Calibration is only as accurate as the test gas used. BW Technologies quality test gases are made to the highest accuracy and trace-ability to N.						
I.S.T. Standards.						
		DE				
Calibrated By: Sara Tse Service Hotline: 2592 2120 Ms. Tse - Service Dept.						
Asia Pacific Industrial Safety Equipment Asia Tackyologias						

Asia Technologies 亞洲科技