

aurecon



# Landfill Gas Monitoring Equipment Calibration Certificate

The copyright of this document is owned by Acuity Sustainability Consulting Limited. It may not be reproduced except with prior written approval from the Company.



香港新界葵涌葵昌路58-70號永祥工業大廈10樓B室

Unit B, 10/F., Wing Cheung Industrial Building, 58-70 Kwai Cheong Road, Kwai Chung, New Territories, HK Tel: (852) 2751 7770 Fax: (852) 2756 2051 E-mail: rotter@rotter.com.hk

### **Calibration Report - Gas Detector**

PGM-2500 (QRAE III) --- LEL/O2/CO/H2S

#### **UNIT INFORMATION :**

| Customer: | Penta Ocean Construction Co Ltd | Serial # : | M02A001708  | Model :    | QRAE III      |
|-----------|---------------------------------|------------|-------------|------------|---------------|
|           |                                 | Firmware : | V2.14       | Sensor :   | LEL/02/CO/H2S |
|           |                                 | Cal date : | 28-Jul-2023 | Inspected: | Teddy         |
|           |                                 | _          |             |            |               |

#### **SENSOR DATA :**

|   | LEL sensor (ME)          | O2 sensor                   | CO sensor (Tox1)    | H2S sensor (Tox2)         |
|---|--------------------------|-----------------------------|---------------------|---------------------------|
| Calibration dates:                                  | 28-Jul-2023              | 28-Jul-2023                 | 28-Jul-2023         | 28-Jul-2023               |
| After Calibration levels                            | 51%                      | 18.10%                      | 50 ppm              | 10.0 ppm                  |
| Alarm levels (Low):                                 | 10.00%                   | 19.50%                      | 35 ppm              | 10 ppm                    |
| Alarm levels (High):                                | 20.00%                   | 23.50%                      | 200 ppm             | 20 ppm                    |
| TWA Level :   |                          |                             | 35 ppm              | 10 ppm                    |
| STEL Level :  |                          |                             | 100 ppm             | 15 ppm                    |
| Status:<br>Pump Speed<br>Clock<br>LEL Gas Selection | Low<br>Yes               | Back Light<br>Measure       | Manual<br>Average   |                           |
| LEL Calibration Gas                                 | Methane                  | LEL measurement Gas         | Methane             |                           |
| LEL Custom Gas                                      | LEL_custom_gas           | LEL Custom Factor           | 1.0                 | -                         |
|   |                          | CO, 10ppm H2S, 50% LE       |                     | Gas lot :302-402538759-74 |
| Fresh Air Calibratio                                | on is nightly recommende | ed to proceed prior for mea | surement each time. |                           |

Replaced Parts:

Notes: The unit was calibrated and checked under good working condition

\*\*Next calibration due on or before 27 July 2024

Teddy Wang Serviced by Rotter International Ltd

# PROMAT (HK) LTD

寶時(香港)有限公司

901 New Trend Centre, 704 Prince Edward Road East, San Po Kong, Kowloon, Hong Kong Tel: (852)2661 2392 Fax: (852)2661 2086 Email:sales@promat.hk http://www.promat.hk



## VERIFICATION CERTIFICATE OF CO2 METER

| Report No. | : 23030                                |
|------------|--|
| Date       | : 27/11/2023                           |
| Client     | : Penta-Ocean-Concentric Joint Venture |

#### **EQUIPMENT TO BE VERIFIED**

| Equipment Name       | : CO2 Meter  |
|----------------------|--------------|
| Supplier             | : TES        |
| Model No.            | : TES-1370H  |
| Serial No.           | : 200901259  |
| Date of Verification | : 17/11/2023 |
| Due Verification     | : 16/11/2024 |

#### VERIFICATION DEVICES USED

| Reference Equipment | : CO2 in N2    |
|---------------------|----------------|
| Supplier            | NorLab         |
| Model No.           | : H1013.5VN    |
| Lot #               | 1-006-21       |
| Expiry date         | : 1/1/2024     |
| Accuracy            | : Within +/-2% |
|                     |                |

#### ENVIRONMENTAL CONDITION

| Ambient Temp      | : | 24.9°C |
|-------------------|---|--------|
| Relative Humidity | : | 50%    |

#### **Verification Result**

| Test Number | Concentration (Mole%) | Results |
|-------------|-----------------------|---------|
| Test 1      | 0.50%                 | 4908ppm |

#### Remarks

1 The Gas reference used in this verification has traceable accuracy to Manufacturer Standard

2 The above equipment was operated by the competent person

3 Promat is Registered ISO9001:2015 Quality Management System in Sales, Repair and Calibration Services

#### Certification

Verification by

Ms. Ning Lee / Service Coordinator

Checked by

Mr. Hei Kong / Technical Engineer

SGS Registered ISO9001:2015 Quality Management System of Sales, Maintenance & Calibration of Testing Equipment