

aurecon



# Landfill Gas Monitoring Equipment Calibration Certificate

The copyright of this document is owned by Acuity Sustainability Consulting Limited. It may not be reproduced except with prior written approval from the Company.



香港新界葵涌葵昌路58-70號永祥工業大廈10樓B室

Unit B, 10/F., Wing Cheung Industrial Building, 58-70 Kwai Cheong Road, Kwai Chung, New Territories, HK Tel: (852) 2751 7770 Fax: (852) 2756 2051 E-mail: rotter@rotter.com.hk

### **Calibration Report - Gas Detector**

PGM-2500 (QRAE III) --- LEL/O2/CO/H2S

#### **UNIT INFORMATION :**

Customer:	Penta Ocean Construction Co Ltd	Serial # :	M02A001708	Model :	QRAE III
		Firmware :	V2.12	Sensor :	LEL/O2/CO/H2S
		Cal date :	28-Jul-2022	Inspected:	Teddy
				]	

#### **SENSOR DATA:**

<u>LEL sensor (ME)</u>	<u>O2 sensor</u>	CO sensor (Tox1)	H2S sensor (Tox2)
28-Jul-2022	28-Jul-2022	28-Jul-2022	28-Jul-2022
50%	18.00%	50 ppm	10.0 ppm
10.00%	19.50%	35 ppm	10 ppm
20.00%	23.50%	200 ppm	20 ppm
		35 ppm	10 ppm
		100 ppm	15 ppm
Yes	Measure	Average	
Methane	LEL measurement Gas	Methane	
LEL_custom_gas	LEL Custom Factor	1.0	
Mix: (18% O2, 50ppm	CO, 10ppm H2S, 50% LE	L CH4, BAL N2)	Gas lot #WO350201-3
	50% 10.00% 20.00%    Low Yes Methane LEL_custom_gas	50%  18.00%    10.00%  19.50%    20.00%  23.50%	50%  18.00%  50 ppm    10.00%  19.50%  35 ppm    20.00%  23.50%  200 ppm      35 ppm      100 ppm      100 ppm      100 ppm    Measure  Manual  Average    Methane  LEL measurement Gas  Methane

**Replaced Parts:** 

Notes: The unit was calibrated and checked under good working condition

\*\*Next calibration due on or before 27 July 2023

Serviced by Teddy Wong



## PROMAT (HK) LTD

寶時(香港)有限公司

901 New Trend Centre, 704 Prince Edward Road East, San Po Kong, Kowloon, HK Tel.: 2661 2392 Fax.: 2661 2086 email : service@promat.hk http://www.promat.hk



Your Solution To Testing Instrument

## VERIFICATION CERTIFICATE OF CO2 Analyzer

Report No.	:	21012
Date	:	15/11/2021
Client	:	Penta Ocean Concentric JV

#### EQUIPMENT TO BE VERIFIED

Equipment Name	:	CO2 Analyzer
Supplier	:	TES
Model No.	:	1307H
Serial No.	:	200901259
Date of Verification	:	15/11/2021
Due Verification	:	14/11/2022

#### VERIFICATION DEVICES USED

Reference Equipment	: CO2 in N2	CO2 in N2
Supplier	: NorLab	NorLab
Model No.	: H1013500PN	H1013.3VN
Lot #	: 0-353-780	1-006-21
Expiry date	: 23/12/2023	10/1/2024
Accuracy	: Within +/-2%	Within +/-2%

#### ENVIRONMENTAL CONDITION

Ambient Temp	:	25.2 °C
Relative Humidity	:	53%

#### **Verification Result**

Test Number	Concentration (Mole%)	Results		
Test 1	500ppm	510ppm		
Test 2	0.50%	0.52%		

#### Remarks

1 The Gas reference used in this verification has traceable accuracy to Manufacturer Standard

2 The above equipment was operated by the competent person

3 Promat is Registered ISO9001:2015 Quality Management System in Sales, Repair and Calibration Services

Certification
Verification by
AC THE THE
Mr. Eric Lee (Product Manager )(1)

	Calil	bration Cei	tificate			
Cert. Ref. No.: BW/XT/3RD/17974 D			ate: 2022 09	02		
			Purchase Order No.: SME-C-20-21-6/2020-76554			
ノレ内E 世元 7년 797頁 2013			ate 2020 07	07 INVOICE	NO: AP	
		E	mail: damonhuang@	renopipe.com.hk		
Attn: Damon Huang	<b>Tel:</b> 3998 3	3193 F	ax: 3998 3225	Mobi	ile Phone	
User Details:						
Gas Detector Model: XT-XW Calibration Record:	HM-Y-OR Seri	ial No.: MA220-0127	09 <b>Pum</b> p	<b>S/N</b> : 420373		
Inpectio	on before calibration	1	Visual insp	ection	Functional Test	
Basic Unit - Case, Clip & Di	splay etc.		OK		ОК	
Battery and charge etc.			OK		ОК	
Motorized Pump			OK		ОК	
Other items						
Turna of Songar			Expiry D	ate		
Type of Sensor Oxygen Sensor						
CO & H2S Sensor						
Combustible(LEL) Sensor						
	Dete of cellineties				LEL (%)	
Type of calibration	Date of calibration	H2S (ppm)	CO (ppm)	O2 (%)		
3rd Calibration	2022 09 02	25	100	18	50	
Result of Cali	bration	OK	ОК	ОК	OK	
Calibration Cost: (As per	attached invoice)	F.O.C				
Calibration remarks: Oxygen Warrant	sensor replaced by new on y : Oxygen Sensor 1 years v					
Next calibration date		<u> </u> _	023 09	02		
USERS MUST READ TH AND FOLLOW THEIR O		NUAL THOROU	GHLY BEFORE O	PERATING 7	THIS EQUIPMENT	
All gas detection instrumer as accurate as the test gas u I.S.T. Standards.						
Calibrated By:	Sara Tse	Service	lotline: 2592 2120	) Ms. Tse - S	Service Dept.	
Asia Pacific Unit B, 1/ł Kv	Industrial Sa		nent		hnologies 科技	