



# Appendix I

## Landfill Gas Monitoring Equipment Calibration Certificate



路達國際有限公司

ROTTER INTERNATIONAL LIMITED

香港新界葵涌葵昌路58-70號永祥工業大廈10樓B室

Unit B, 10/F., Wing Cheung Industrial Building, 58-70 Kwai Cheong Road, Kwai Chung, New Territories, HK.

Tel: (852) 2751 7770 Fax: (852) 2756 2051 E-mail: rotter@rotter.com.hk

## Calibration Report - Gas Detector

PGM-2500 (QRAE III) --- LEL/O2/CO/H2S

### UNIT INFORMATION :

Customer:	Penta Ocean Construction Co Ltd	Serial #:	M02A001708	Model:	QRAE III
		Firmware:	V2.12	Sensor:	LEL/O2/CO/H2S
		Cal date:	28-Jul-2021	Inspected:	Teddy

### SENSOR DATA :

	LEL sensor (ME)	O2 sensor	CO sensor (Tox1)	H2S sensor (Tox2)
Calibration dates:	28-Jul-2021	28-Jul-2021	28-Jul-2021	28-Jul-2021
After Calibration levels	50%	17.90%	50 ppm	10.1 ppm
Alarm levels (Low):	10.00%	19.50%	35 ppm	10 ppm
Alarm levels (High):	20.00%	23.50%	200 ppm	20 ppm
TWA Level:	--	--	35 ppm	10 ppm
STEL Level:	--	--	100 ppm	15 ppm

### Status:

Pump Speed	Low	Back Light	Manual
Clock	Yes	Measure	Average

### LEL Gas Selection

LEL Calibration Gas	Methane	LEL measurement Gas	Methane
LEL Custom Gas	LEL_custom_gas	LEL Custom Factor	1.0

Gas types used : 4-Gas Mix: (18% O2, 50ppm CO, 10ppm H2S, 50% LEL CH4, BAL N2) Gas lot #1412983 Cyl# 15

\*\*\* Fresh Air Calibration is highly recommended to proceed prior for measurement each time.

Replaced Parts:

### Notes:

The unit was calibrated and checked under good working condition

\*\*Next calibration due on or before 27 July 2022

Serviced by  Teddy Wong  
Rotter International Ltd

# PROMAT (HK) LTD

寶時(香港)有限公司

901 New Trend Centre, 704 Prince Edward Road East, San Po Kong, Kowloon, HK  
Tel.: 2661 2392 Fax.: 2661 2086 email : service@promat.hk http://www.promat.hk



Your Solution To Testing Instrument

## VERIFICATION CERTIFICATE OF CO2 Analyzer

Report No. : 21012  
Date : 15/11/2021  
Client : Penta Ocean Concentric JV

### EQUIPMENT TO BE VERIFIED

Equipment Name : CO2 Analyzer  
Supplier : TES  
Model No. : 1307H  
Serial No. : 200901259  
Date of Verification : 15/11/2021  
Due Verification : 14/11/2022

### VERIFICATION DEVICES USED

Reference Equipment	: CO2 in N2	CO2 in N2
Supplier	: NorLab	NorLab
Model No.	: H1013500PN	H1013.3VN
Lot #	: 0-353-780	1-006-21
Expiry date	: 23/12/2023	10/1/2024
Accuracy	: Within +/-2%	Within +/-2%

### ENVIRONMENTAL CONDITION

Ambient Temp : 25.2 °C  
Relative Humidity : 53%

### Verification Result


Test Number	Concentration (Mole%)	Results
Test 1	500ppm	510ppm
Test 2	0.50%	0.52%

### Remarks

- 1 The Gas reference used in this verification has traceable accuracy to Manufacturer Standard
- 2 The above equipment was operated by the competent person
- 3 Promat is Registered ISO9001:2015 Quality Management System in Sales, Repair and Calibration Services

### Certification

Verification by

  
Mr. Eric Lee / Product Manager

